

## North Krios 165k× NonCDS, Pixel 0.51099Å, 4 Shots/Hole

### LMM Settings

- Spot Size 6, C2 70μm, EFTEM LM 175×, Ill Area bigger than detector, Default Montage Settings

### MMM Settings

- Spot Size 6, C2 70μm, EFTEM LM 580×, Ill Area bigger than detector

### Low Dose Settings

- Spot Size 6, C2 70μm
- View - EFTEM, Nanoprobe, SA 15000×
- R/F/T - EFTEM, Nanoprobe, SA 165k×, Ill Area 480nm
- EPS ~15

### Gain Ref Path

- C: ProgramData>Gatan>Reference Images>Most recent and Largest File size ~360,000KB
- File Format .DM4
- Relion Processing - Flip left to right OR rotate 180 and flip Y
- CryoSPARC Processing - Flip in both X and Y